

TEST FACILITY

NAMSA
 115 Chemin de L'Ision
 38670 Chasse-sur-Rhône
 France
 Tel : +33 4 78 07 92 34
 Fax : +33 4 72 24 08 12

SPONSOR

PLATIT AG
 Dr. Hamid Bolvardi
 Eichholzstr. 9
 2545 Selzach / SO
 Switzerland

CONFIDENTIAL

Test Article Name	TiCN PVD coated Ti6Al4V plates
Article Identification	Reference: Biocompatible uncoated Ti6Al4V material, Batch: TiCN

Test	Standard	Study number	Results
ISO MTS cytotoxicity test	ISO 10993-5 (2009)	292762	Pass*
ISO Intracutaneous Study in Rabbits (Two Extracts)	ISO 10993-10 (2010)	292763	Pass*


*: Conclusions apply only to the test article as received and tested.

Only the final reports are legally acceptable.

Approved by:



 Alizé Laffitte
 Associate Study Director



 Date