

**TEST FACILITY**

NAMSA  
 115 Chemin de L'Ision  
 38670 Chasse-sur-Rhône  
 France  
 Tel : +33 4 78 07 92 34  
 Fax : +33 4 72 24 08 12

**SPONSOR**

**PLATIT AG**  
 Dr. Hamid Bolvardi  
 Eichholzstr. 9  
 2545 Selzach / SO  
 Switzerland

CONFIDENTIAL

Test Article Name	TiN PVD coated Ti6Al4V plates
Article Identification	Reference: Biocompatible uncoated Ti6Al4V material, Batch: TiN

Test	Standard	Study number	Results
ISO MTS cytotoxicity test	ISO 10993-5 (2009)	292770	Pass*
ISO Intracutaneous Study in Rabbits (Two Extracts)	ISO 10993-10 (2010)	292771	Pass*

\*: Conclusions apply only to the test article as received and tested.

Only the final reports are legally acceptable.

Approved by:



\_\_\_\_\_  
 Claire Nahon  
 Associate Study Director

August 05, 2021  
 Date